



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Masterleo, LLC

975 Eastwind Dr #150

Westerville, OH 43081

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to be 'Jason Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 06 May 2028

Certificate Number: AC-3941



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Masterleo, LLC
 975 Eastwind Dr, #150
 Westerville, OH 43081
 Adam Dupps 800-779-1444

CALIBRATION

ISO/IEC 17025 Accreditation Granted: **30 April 2026**

Certificate Number: **AC-3941** Certificate Expiry Date: **06 May 2028**

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Pressure Gages ¹	(-15 to 30) psi	0.009 4 psi	Direct Comparison to Certified Reference Gauges
	(0 to 30) psi	0.04 psi	
	(0 to 100) psi	0.12 psi	
	(0 to 300) psi	0.18 psi	


Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
RTD/TC Thermometers & Transmitters ¹	(-50 to 0) °C	0.5 °C	Direct Comparison to Certified Reference Thermometer using a Drywell or Oil Bath
	(> 0 to 100) °C	0.61 °C	
	(> 100 to 160) °C	0.63 °C	

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

- On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.



Jason Stine, Vice President

This Scope of Accreditation, version 003, was last updated on: 30 April 2026 and is valid only when accompanied by the Certificate.

Page 1 of 1